

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

## CB TEST CERTIFICATE

Product

Switching power supply

Name and address of the applicant

TRACO ELECTRONIC AG  
SIHLBRUGGSTRASSE 111 CH-6340 BAAR,  
SWITZERLAND

Name and address of the manufacturer

TRACO ELECTRONIC AG  
SIHLBRUGGSTRASSE 111 CH-6340 BAAR,  
SWITZERLAND

Name and address of the factory

TRACO ELECTRONIC AG  
SIHLBRUGGSTRASSE 111 CH-6340 BAAR,  
SWITZERLAND

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

 Additional Information on page 2Input: 100-240 Vac, 50/60 Hz,  
1.6 A max. for TPP100-XYzzzzzzzz series,  
2.0 A max. for TPP150-XYzzzzzzzz series.

Output Rating: See test report for details.

Trademark (if any)

TRACO



Type of Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

TPP100-XYzzzzzzzz, TPP150-XYzzzzzzzz  
See Page 2

Additional information (if necessary may also be reported on page 2)

Additionally evaluated to EN 60601-1: 2006 / A1: 2013 / A12: 2014;

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

As shown in the Test Report Ref. No. which forms part of this Certificate

170102104 issued on 2017-03-30

This CB Test Certificate is issued by the National Certification Body



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA



UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK



UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN



UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2017-04-03

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

**DK-62477-UL**

**Model Details:**

TPP100-XYzzzzzzzz (X can be 112, 115, 124, 128, 136 and 148 or any other number within SELV limit; Y can be A, U, D or blank; z can be any alphanumeric or blank for marketing purpose and no impact safety related critical components and constructions)

TPP150-XYzzzzzzzz (X can be 105, 112, 115, 124, 128, 136 and 148 or any other number within SELV limit; Y can be A, U, D or blank; z can be any alphanumeric or blank for marketing purpose and no impact safety related critical components and constructions)

**Additional information (if necessary)**



- 
- 
- 
- 

UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2017-04-03

Signature:

Jan-Erik Storgaard