



Ref. Certif. No.

US-19462-ULIEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC**Product
Produit

Built-in type Power Supply

Name and address of the applicant
Nom et adresse du demandeurName and address of the manufacturer
Nom et adresse du fabricantTRACO ELECTRONIC AG
JENATSCHSTRASSE 1
CH-8002 ZURICH, SWITZERLANDName and address of the factory
Nom et adresse de l'usine

Note: When more than one factory, please report on page 2
Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2^{ème} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales☒ Additional Information on page 2
See Page 2Trademark (if any)
Marque de fabrique (si elle existe)Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeurModel / Type Ref.
Ref. De typeTMT15105, TMT15112, TMT15115, TMT15124, TMT15212,
TMT15215, See Page 2

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2^{ème} page

Risk Management was not included in this investigation.
National Differences specified in the CB Test Report.

☐ Additional Information on page 2

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1(ed.3)

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

E141988-A65-CB-1 issued on 2012-07-31

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- ☒ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☐ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2012-07-31

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-19462-UL

Model Details:

TMT15105C, TMT15112C, TMT15115C, TMT15124C, TMT15212C,
TMT15215C

Factories:



Ratings:

For Models TMT15105, TMT15105C,
Input: 115-230V~, 50/60Hz, 0.29-0.19A,
Output: 5VDC, 3000mA

For Models TMT15112, TMT15112C,
Input: 115-230V~, 50/60Hz, 0.29-0.19A,
Output: 12VDC, 1250mA

For Models TMT15115, TMT15115C,
Input: 115-230V~, 50/60Hz, 0.29-0.19A,
Output: 15VDC, 1000mA

For Models TMT15124, TMT15124C,
Input: 115-230V~, 50/60Hz, 0.29-0.19A,
Output: 24VDC, 625mA

For Models TMT15212, TMT15212C,
Input: 115-230V~, 50/60Hz, 0.28-0.18A,
Output: ± 12 VDC, 625mA

For Models TMT15215, TMT15215C,
Input: 115-230V~, 50/60Hz, 0.28-0.18A,
Output: ± 15 VDC, 500Ma

Additional information (if necessary)

Information complémentaire (si nécessaire)



☒ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

☐ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2012-07-31

Signature:

Jolanta M. Wroblewska