

Traco Electronic Co. Ltd. Sihlbruggstrasse 111 6340 Baar

T +41 43 311 45 11 www.tracopower.com For technical requests: www.tracoserviceportal.com



Traco Electronic Co. Ltd. Sihlbruggstrasse 111 6340 Baar

T +41 43 311 45 11 www.tracopower.com For technical requests: www.tracoserviceportal.com

DECLARATION OF CONFORMITY



Function Module, TSP-BCMU Series

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: TSP-BCMU360(z) (z) can be blank, any number of alphanumeric characters, "-", or punctuation marks LVD: 2014/35/EU EMC: 2014/30/EU RoHS: 2011/65/EU

EN 61010-1:2010 + A1:2019 Safety standards: EN 61010-2-201:2018

EMC standards: EN 61000-6-2:2005 + AC:2005

EN 61000-6-3:2007 + A1:2011 + AC:2012

EN 55032:2012

Year of CE marking:

2011

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.

J. Jee Harry.

DECLARATION OF CONFORMITY Function Module, TSP-BCMU Series

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models:

TSP-BCMU360(z)

(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks

Electrical Equipment (Safety) Regulations 2016

Electromagnetic Compatibility Regulations 2016

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

Safety standards:

EN 61010-1:2010 + A1:2019

EN 61010-2-201:2018

EMC standards:

EN 61000-6-2:2005 + AC:2005

EN 61000-6-3:2007 + A1:2011 + AC:2012

EN 55032:2012

Year of UKCA marking:

Baar, February 21, 2024

2021

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.

J. Jee Harris.